SUPPORTING INFECTIOUS DISEASE RESEARCH

Staphylococcus haemolyticus, Strain NRS62

Catalog No. NR-45892

Product Description: *Staphylococcus haemolyticus (S. haemolyticus)*, strain NRS62 was isolated in 2000 from the dialysis fluid of a 58-year-old female renal ward inpatient in Texas, USA. *S. haemolyticus*, strain NRS62 was deposited as a glycopeptide-intermediate *S. haemolyticus* strain.

Lot¹: 70011777

Manufacturing Date: 12JAN2018

TEST	SPECIFICATIONS	RESULTS
Phenotypic Analysis		
Cellular morphology	Gram-positive cocci	Gram-positive cocci
Colony morphology ²	Report results	Circular, convex, entire, smooth and
		white (Figure 1)
Motility (wet mount)	Report results	Non-motile
Hemolysis ²	Report results	β-hemolytic
Biochemical characterization		
Catalase	Positive	Positive
Coagulase ³	Negative	Negative
VITEK [®] 2 Compact (GP card)	S. haemolyticus (≥ 89%)	S. haemolyticus (99%)
Antibiotic Susceptibility Profile ⁴		
VITEK [®] (AST-GP71 card)		
Beta-lactamase ⁵	Report results	Positive
Cefoxitin screen	Report results	Negative ⁶
Benzylpenicillin	Report results	Resistant (≥ 0.5 µg/mL)
Oxacillin	Resistant	Resistant (= 1-2 µg/mL)
Gentamicin	Resistant	Resistant (≥ 16 µg/mL)
Ciprofloxacin	Resistant	Resistant (≥ 8 µg/mL)
Levofloxacin	Report results	Resistant (≥ 8 µg/mL)
Moxifloxacin	Report results	Resistant (= 4 µg/mL)
Clindamycin (inducible resistance)	Report results	Positive ⁷
Erythromycin	Resistant	Resistant (≥ 8 µg/mL)
Quinupristin/Dalfopristin	Sensitive	Sensitive (= 0.5 µg/mL)
Linezolid	Sensitive	Sensitive (= 2 µg/mL)
Daptomycin	Susceptible	Susceptible (= 1 µg/mL)
Minocycline	Report results	Sensitive (≤ 0.5 µg/mL)
Tetracycline	Report results	Resistant (≥ 16 µg/mL)
Tigecycline	Report results	Resistant (= 1 μ g/mL) ⁸
Nitrofurantoin	Report results	Sensitive (≤ 16 µg/mL)
Rifampicin	Report results	Sensitive (≤ 0.5 µg/mL)
Trimethoprim/sulfamethoxazole	Resistant	Resistant (= $80 \mu g/mL$)
Etest [®] antibiotic test strips ⁹		
Teicoplanin	Resistant	Intermediate (= 24 µg/mL) ¹⁰
Vancomycin	Intermediate	Intermediate (= 12-16 µg/mL)
Genotypic Analysis		
Sequencing of 16S ribosomal RNA gene	\geq 99% sequence identity to	99.9% sequence identity to
(~ 750 base pairs)	<i>S. haemolyticus</i> , type strain	<i>S. haemolyticus,</i> type strain
(· · · · · · · · · · · · · · · · · · ·	(GenBank: D83367.1)	(GenBank: D83367.1) ¹¹
		· · · · ·
Purity (post-freeze) ¹²	Consistent with expected colony	Consistent with expected colony
· ·	morphology	morphology
Viability (post-freeze) ²	Growth	Growth
	REL Resources as part of the NARSA collection	

¹S. haemolyticus, strain NRS62 was deposited to BEI Resources as part of the NARSA collection. NR-45892 was produced by inoculation of the deposited material into Tryptic Soy broth and grown 1 day at 37°C in an aerobic atmosphere. Broth inoculum was added to Tryptic Soy agar with 5% defibrinated sheep blood kolles which were grown 1 day at 37°C in an aerobic atmosphere to produce this lot.

²1 day at 37°C in an aerobic atmosphere on Tryptic Soy agar with 5% defibrinated sheep blood ³1 day at 37°C in rabbit serum with 0.15% EDTA (Coagulase Plasma BBL™ 240827)

BEI Resources www.beiresources.org E-mail: <u>contact@beiresources.org</u> Tel: 800-359-7370 Fax: 703-365-2898 **b**|**e**|**i** resources

SUPPORTING INFECTIOUS DISEASE RESEARCH

⁴Minimum Inhibitory Concentration (MIC); MIC Interpretation Guideline: CLSI M100-S22 (2012)

⁵The production of beta-lactamase was detected using a Cefinase™ Paper Disc (BBL™ 231650).

- ⁶The VITEK® AST-GP71 card tests both cefoxitin and oxacillin susceptibilities to determine the final oxacillin interpretation based on CLSI guidelines. However, for coagulase-negative *Staphylcoccal* (CoNS) species, it has been reported that VITEK cards can fail to detect cefoxitin resistance in *mecA* positive strains. Please refer to Johnson, K. N., K. Andreacchio and P. H. Edelstein. "Detection of Methicillin-Resistant Coagulase-Negative *Staphylcoccci* by the VITEK[®] 2 System." <u>J. Clin. Microbiol.</u> 52 (2014): 3196-3199 PubMed: 1500504 for additional information about detection of methicillin-resistant CoNS strains.
- ⁷The VITEK[®] AST-GP71 card tests for both clindamycin resistance and inducible clindamycin resistance (ICR). A positive ICR test is indicative of inducible MLS_b resistance, which confers resistance to macrolides, lincosamides, and type B streptogramin and the isolate should be considered resistant to clindamycin. S. *haemolyticus*, strain NRS62 was found to be sensitive to clindamycin but had a positive ICR test and therefore is considered resistant to clindamycin. For additional information, please refer to Mahesh, C. B., B. K. Ramakant and V. S. Jagadeesh. "The Prevalence of Inducible and Constitutive Clindamycin Resistance Among the Nasal Isolates of *Staphylococci.*" J. Clin. Diagn. Res. 7 (2013): 1620-1622. PubMed: 24086856.
- ⁸MIC Interpretation Guideline: EUCAST Version 4.0 (2014)

⁹1 day at 37°C in an aerobic atmosphere on Mueller Hinton agar

¹⁰S. haemolyticus, strain NRS62 was deposited as being resistant to teicoplanin. Antibiotic susceptibility testing using bioMérieux Etest® antibiotic test strips and performed in duplicate determined that strain NRS62 has intermediate susceptibility to teiocoplanin. For additional information, please refer to Walsh, T. R., et al. "Evaluation of Current Methods for Detection of Staphylococci with Reduced Susceptibility to Glycopeptides." J. Clin. Microbiol. 39 (2001): 2439-2444. PubMed: 11427551.

¹¹Also consistent with other *Staphylococcus* species

¹²Purity of this lot was assessed for 7 days at 37°C in an aerobic atmosphere with 5% CO₂ on Tryptic Soy agar with 5% defibrinated sheep blood.




Figure 1: Colony Morphology

05 JUN 2018

Program Manager or designee, ATCC Federal Solutions

ATCC[®] is a trademark of the American Type Culture Collection.

You are authorized to use this product for research use only. It is not intended for human use.

ATCC[®], on behalf of BEI Resources, hereby represents and warrants that the material provided under this certificate has been subjected to the tests and procedures specified and that the results described, along with any other data provided in this certificate, are true and accurate to the best of ATCC[®]'s knowledge.

(