

Bacteriophage Phi05_1999M/Ec2424, Infectious for *Escherichia coli*

Catalog No. HM-672

Product Description: Bacteriophage phi05_1999M/Ec2424 was derived from a blood sample taken from a patient with septicemia.

Lot^{1,2}: 60185060

Manufacturing Date: 15SEP2011

TEST	SPECIFICATIONS	RESULTS
Plaque Morphology ³	Report results	Circular and clear
Titer ³ Plaque-forming units (pfu) with <i>Escherichia coli</i>	>10 ⁶ pfu per mL	8 x 10 ⁹ pfu per mL
Sterility	0.22 µm sterile-filtered	0.22 µm sterile-filtered

¹Quality control of HMP material is only performed to demonstrate that the product distributed by BEI Resources is identical to the deposited material. It should not be considered a complete characterization of the deposited material.

²The deposited material was added to a Luria-Burtani (LB) soft agar overlay (0.5%) supplemented with 10 mM MgSO₄ containing the phage host, *Escherichia coli*, strain Ec2424. The soft agar overlay was added to LB kolles supplemented with 10 mM MgSO₄ and incubated 24 hours at 30°C in an aerobic atmosphere. Three additional passages occurred under the above conditions. Bacteriophage was harvested with LB broth after each passage, filtered and pooled to produce this lot.

³Lysis was observed on the soft agar overlay after 24 hours at 30°C in an aerobic atmosphere.

Date: 07 MAY 2014

Signature: 

Title: Technical Manager, BEI Authentication or designee

ATCC[®], on behalf of BEI Resources, hereby represents and warrants that the material provided under this certificate has been subjected to the tests and procedures specified and that the results described, along with any other data provided in this certificate, are true and accurate to the best of ATCC[®]'s knowledge.

ATCC[®] is a trademark of the American Type Culture Collection.

You are authorized to use this product for research use only. It is not intended for human use.

