

Staphylococcus aureus, Strain RN0451

Catalog No. NR-45938

Product Description: *Staphylococcus aureus* (*S. aureus*), strain RN0451 is lysogenic for phage Φ11 and was derived from *S. aureus*, strain RN0450 (NRS135). In turn, strain RN0450 was derived from successive UV treatment of *S. aureus*, strain NCTC8325 (NRS77).


Lot¹: 62471632

Manufacturing Date: 02APR2014

TEST	SPECIFICATIONS	RESULTS
Phenotypic Analysis Cellular morphology ² Colony morphology ² Motility (wet mount) Hemolysis ² Biochemical characterization Catalase Coagulase ³ VITEK [®] 2 Compact (GP card) VITEK [®] MS (MALDI-TOF)	Gram-positive cocci Report results Report results Report results Positive Report results ≥ 90% probability of being <i>S. aureus</i> Consistent with <i>S. aureus</i>	Gram-positive cocci Circular, low convex, entire, smooth and gray (Figure 1) Non-motile β-hemolytic Positive Positive <i>S. aureus</i> (99% probability) ⁴ <i>S. aureus</i> (99.9%)
Antibiotic Susceptibility Profile VITEK [®] (AST-GP71card) ⁵ Beta-lactamase ⁶ Cefoxitin screen Benzylpenicillin Oxacillin Gentamicin Ciprofloxacin Levofloxacin Moxifloxacin Clindamycin (inducible resistance) Erythromycin Clindamycin Quinupristin/dalfopristin Linezolid Daptomycin Vancomycin Minocycline Tetracycline Tigecycline Nitrofurantoin Rifampicin Trimethoprim/sulfamethoxazole Etest [®] antibiotic test strips ⁸ Chloramphenicol ⁹ Teicoplanin ⁹	Report results Report results Report results Sensitive Sensitive Sensitive Report results Report results Report results Sensitive Sensitive Sensitive Sensitive Report results Sensitive Report results Report results Report results Report results Report results Sensitive Report results Sensitive	Negative Negative Sensitive (≤ 0.03 µg/mL) Sensitive (≤ 0.25 µg/mL) Sensitive (≤ 0.5 µg/mL) Sensitive (≤ 0.5 µg/mL) Sensitive (≤ 0.12 µg/mL) Sensitive (≤ 0.25 µg/mL) Negative Sensitive (≤ 0.25 µg/mL) Sensitive (≤ 0.25 µg/mL) Sensitive (≤ 0.25 µg/mL) Sensitive (= 2 µg/mL) Sensitive (= 0.25 µg/mL) Sensitive (= 1 µg/mL) Sensitive (≤ 0.5 µg/mL) Sensitive (≤ 1 µg/mL) Sensitive (≤ 0.12 µg/mL) ⁷ Sensitive (≤ 16 µg/mL) Sensitive (≤ 0.5 µg/mL) Sensitive (≤ 10 µg/mL) Sensitive (= 1 µg/mL) Sensitive (= 1 µg/mL)
Genotypic Analysis Sequencing of 16S ribosomal RNA gene (~ 1450 base pairs)	≥ 99% sequence identity to <i>S. aureus</i> type strain (GenBank: L37597)	99.9% sequence identity to <i>S. aureus</i> type strain (GenBank: L37597)
Purity (post-freeze)¹⁰	Consistent with expected colony morphology	Consistent with expected colony morphology
Viability (post-freeze)²	Growth	Growth

- ¹*S. aureus*, strain RN0451 was deposited to BEI Resources as part of the NARSA collection. NR-45938 was produced by inoculation of the deposited material into Tryptic Soy broth and grown 23 hours at 37°C in an aerobic atmosphere. Broth inoculum was added to Tryptic Soy agar with 5% defibrinated sheep blood kolles which were grown 24 hours at 37°C in an aerobic atmosphere to produce this lot.
- ²24 hours at 37°C in an aerobic atmosphere on Tryptic Soy agar with 5% defibrinated sheep blood
- ³24 hours at 37°C in rabbit serum with 0.15% EDTA (Coagulase Plasma BBL™ 240827)
- ⁴Percent probabilities above 90% indicate a close match to the typical biochemical pattern for the given organism, with a percent probability of 99% being a perfect match between the test reaction pattern and the unique biochemical pattern of the given organism or organism group. For additional information, please refer to O'Hara, C.M. and J. M. Miller. "Evaluation of the Vitek 2 ID-GNB Assay for Identification of Members of the Family *Enterobacteriaceae* and Other Nonenteric Gram-Negative Bacilli and Comparison with the Vitek GNI+ Card." *J. Clin. Microbiol.* 41 (2003): 2096-2101. PubMed: 12734254.
- ⁵Minimum Inhibitory Concentration (MIC); MIC Interpretation Guideline: CLSI M100-S22 (2012)
- ⁶The production of beta-lactamase was detected using a Cefinase™ Paper Disc (BBL™ 231650).
- ⁷MIC Interpretation Guideline: EUCAST Version 4.0 (2014)
- ⁸24 hours at 37°C in an aerobic atmosphere on Mueller Hinton agar
- ⁹For both chloramphenicol (bioMérieux Etest® 412308) and teicoplanin (bioMérieux Etest® 412459), a MIC ≤ 8 µg/mL is sensitive, a MIC = 16 µg/mL is intermediate, and a MIC ≥ 32 µg/mL is resistant.
- ¹⁰Purity of this lot was assessed for 7 days at 37°C in an aerobic atmosphere on Tryptic Soy agar with 5% defibrinated sheep blood.

Figure 1: Colony Morphology



Date: 16 JUL 2014

Signature:

BEI Resources Authentication

ATCC®, on behalf of BEI Resources, hereby represents and warrants that the material provided under this certificate has been subjected to the tests and procedures specified and that the results described, along with any other data provided in this certificate, are true and accurate to the best of ATCC®'s knowledge.

ATCC® is a trademark of the American Type Culture Collection.

You are authorized to use this product for research use only. It is not intended for human use.

