

⁴Minimum Inhibitory Concentration (MIC); MIC Interpretation Guideline: CLSI M100-S22 (2012)

⁵MIC interpretation was determined using VITEK[®] 2 software version 07.01 combined with the bioMérieux Advanced Expert System™ (AES) software using the interpretation standard CLSI M100-S22 (2012) and the interpretation guideline "Natural Resistance." For more information, please refer to Sanders, C. C., et al. "Potential Impact of the VITEK 2 System and the Advanced Expert System on the Clinical Laboratory of a University-Based Hospital." *J. Clin. Microbiol.* 39 (2001): 2379-2385. PubMed: 11427542.

⁶1 day at 37°C in an aerobic atmosphere on Mueller Hinton agar

⁷Also consistent with other organisms

⁸Purity of this lot was assessed for 7 days at 37°C in an aerobic atmosphere with 5% CO₂ on Tryptic Soy agar with 5% defibrinated sheep blood.

Figure 1: Colony Morphology



Date: 14 NOV 2017

Signature:

BEI Resources Authentication

ATCC[®], on behalf of BEI Resources, hereby represents and warrants that the material provided under this certificate has been subjected to the tests and procedures specified and that the results described, along with any other data provided in this certificate, are true and accurate to the best of ATCC[®]'s knowledge.

ATCC[®] is a trademark of the American Type Culture Collection.

You are authorized to use this product for research use only. It is not intended for human use.

